

EN • MINI2 FX – DEVICE

Device for aesthetic use. Does not perform diagnosis and treatments. CE mark.

Cod. 126212

Manufacturer

Callegari S.r.l.

Description and Intended Use

Mini2 FX is a device designed for aesthetic purposes, to determine the melanin content of the skin.

Additionally, if desired, it is possible to estimate the phototype according to the Fitzpatrick scale.

The device is intended for use in pharmacies, parapharmacies, beauty salons and clinics, cosmetic companies and in any environment where skin analysis for purely aesthetic purposes is desired.

Mini2 FX is a device for aesthetic use. It does not perform diagnosis or treatments of any kind.

Technical features

Interface	TFT graphic colour display 0.96" + 1 button for measurement activation
Power Supply	2 x 1.5V AAA batteries (not provided)
Exterior appearance	Pen-type probe in ABS with tip sensor
Weight	50 g (excluding batteries)
Dimensions	160 (w) x 33(d) x 25(h) mm
Operating conditions	Temperature -18 °C- +55 °C; max. relative humidity 90% not-condensed
Storage conditions	Temperature -30 °C - +80°C; max. relative humidity 90% not-condensed (excluding batteries).
	Temperature - 15 °C- +35 °C max relative humidity 60% not-condensed (including batteries)
Parameter	Melanin
Measurance Principle	Double wavelength reflectance photometer ($\lambda_1=875\text{nm}$, $\lambda_2=660\text{nm}$)
Range	0 – 100 a. u. (arbitrary units)
Resolution	1 a. u.
Compliance	CE mark